applied biosystems
by Thermo Fisher Scientific

Certificate of Analysis

Capillary Array 36cm 3500 Series

REF 4404683

SN L324J6415

30-Oct-2024

25-Sep-2025

TEST	SPECIFICATION	RESULT
Capillary Leak Test Each array is tested to assure absence of leaks.		
Leak Test	No Leaks	PASS
Capillary Plug Test Each array is tested to assure absence of plugged capillaries.		
Plug Test	No Plugged Capillaries	PASS
Fluorescence Background Test Each array is tested to assure background fluorescence is within specification.		
Fluorescence Background Test	Conforms to Specification	PASS
Capillary Mapping Test Each array is tested to verify that capillary inlet and outlet spatial positioning correspond.		
Mapping Test	Conforms to Specification	PASS

For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.

ISO13485 REGISTERED FACILITY

Singapore

Electronically signed by: Mei Ying Vivian Khoo
Reason: Reviewer of the GxP document
Date: Apr 17, 2025 15:24 GMT+8

Quality Assurance

(Name, Signature & Date)

Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06, Singapore 739256 Tel: (65) 6362 9300
To obtain a Certificate of Analysis on-line go to www.thermofisher.com or email us at CofARequests@thermofisher.com