Certificate of Analysis

ThermoFisher SCIENTIFIC

Cathode Buffer Container 3500 Series

Product No.	**4408256**
Lot No.	**2502066**
Date of Manufacture	**18MAR2025**
Expiration Date	**20MAY2026**

TEST	SPECIFICATION	RESULT
VISUAL INSPECTION		
Clear, particulate free, viscous solution		Pass
pH TEST Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.876
CONDUCTIVITY Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1767 µS/cm

For Research Use Only.
Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Bedford, MA, USA

Kim Khear
Quality Assurance
Issued 01APR2025

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com