ThermoFisher SCIENTIFIC

DECLARATION OF CONFORMITY

ProFlex™ 3x32-Well PCR System, ProFlex™ 96-Well PCR System,
ProFlex™ Dual 96-Well PCR System, ProFlex™ Dual 384-Well PCR System,
ProFlex™ Dual Flat PCR System, ProFlex™ Base, ProFlex™ 3x32-Well Sample Block,
ProFlex™ 96-Well Sample Block, ProFlex™ Dual 96-Well Sample Block,
ProFlex™ Dual 384-Well Sample Block, ProFlex™ Dual Flat Sample Block

Part Number(s): 4484073, 4484075, 4484076, 4484077, 4484078, 4483636, 4483638, 4483637, 4484071, 4484072, 4484074
Issued: 16 Nov 2017

We declare that, to the best of our knowledge, the products specified above are in conformity with the applicable provisions of the Directives and Standards listed below, and are entitled to bear the CE Mark.

DIRECTIVES	**Low Voltage Directive 2014/35/EU**	**EMC Directive 2014/30/EU**	**RoHS Directive 2011/65/EU**
SUPPORTING STANDARDS	EN 61010-1:2010 EN 61010-2-010:2014 EN 61010-2-081:2015	EN 61326-1:2013 Class B EN 61000-3-2:2006+A2:2009 EN 61000-3-3:2008	EN 50581:2012

MANUFACTURER & FACTORY
Life Technologies Holdings Pte Ltd.
Blk 33, #07-06, Marsiling Industrial Estate Rd 3
Singapore 739256, Singapore

EC AUTHORIZED REPRESENTATIVE
Life Technologies Ltd
3 Fountain Drive
Inchinnan Business Park
Paisley PA4 9RF
U.K.

Kee Yang Hooi
Staff Engineer,
Compliance Engineering/
Manufacturing Engineering